

第86回(一社)日本脳神経外科学会中国四国支部学術集会、松山、Dec. 7, 2018
会長:國枝武治教授(愛媛大学)、座長:川西正彦(香川大学)、コメンテーター:岡崎貴仁(広島大学)

ASL(Silent)-MRAとFSE-MRCを組合せた 3D-multifusion imageによるCoil治療後 脳動脈瘤のneck/domeの画像評価

・佐藤 透1, 菱川 朋人2, 平松 匠文2,

・杉生 憲志2, 伊達 勲2, 大田 慎三3

・1佐藤脳神経外科, 2岡山大学大学院 脳神経外科,

3脳神経センター大田記念病院 脳神経外科

これから始まります。

Lecturers report no conflict of interest concerning the materials or methods used in this study..



Arterial Spin-Labeling (ASL) とは？ ～動脈血プロトンの反転spinがtracerとなる～

①総頸動脈高位で、Long RF inversion pulse. (**Labeling**)


②頭頂部で Reverse Labeling pulseを収集. (**Control**)

③**Labeling**から**Control**のdatasetを subtractして、T1WI画像化.

● **ASL-perfusion:** 血液spinを内因性 tracerとした非造影脳灌流画像.
CT・MR-perfusion(造影)・SPECT(RI)

● **ASL-MRA:** 磁化率アーチファクトのない管腔内動脈血(血管)形態を描出.



- Signa Pioneer 3.0T(GE)
- Discovery MR750W 3.0T (GE)





ASL(Silent)-MRA Silenz Scan Sequence

Silent Scan Technology

Sequence chart of common 3D data acquisition



Sequence chart of SILENT SCAN (Silenz)



ASL(Silent)-MRA (GE SIGNA Pioneer:3.0T)



TOF vs ASL(Silent)-MRA Scan Parameters

	TOF-MRA	ASL-MRA
原理	in-flow effect	ASL+Ultra short TE+T1WI subtraction
環境騒音	Noisy: >100 dB	Silent: < 3dB
飽和効果	Difficult: slow flow Signal loss : fast, whirlpool, turbulence flow Difficult: parallel running flow	Easy: slow flow Depict: slow, whirlpool, turbulence flow Multi-directional flow
TR/TE	25/3.4	880/0.016
FOV	200 x 200 mm	200 x 200 mm
Matrix	288 x 192	
スローク数		320
スライス厚	1.2 mm	1.0 mm
NEX	0.85	1
撮像時間	4 min 59s	12 min 48s




Cases with Coiling Treatment


Case	Age/Sex	Location	Symptoms	Time to MRA (mo) *
1	63/F	AComA	unruptured	6
2	67/M	AComA	unruptured	3
3	58/M	AComA	unruptured	12
4	62/F	AComA	unruptured	12
5	53/M	AComA	unruptured	3
6	78/F	A2-A3	unruptured	3
7	75/F	Lt ICPC	unruptured	12
8	51/M	IC AChor	unruptured	12
9	56/F	Lt MCA	SAH	12
10	69/F	BA-top	unruptured	3
11	69/F	BA-top	SAH	12
12	53/F	BA-top	unruptured	12



Case 10: BA-top Aneurysm



Case 6: A2–A3 Aneurysm




Size & Volume of Remnants & Dome

Cases	Remnant size (mm) Silent MRA	Remnant volume (mm ³) Silent MRA	Dome size (mm) FSE MRC	Dome volume (mm ³) FSE MRC	Remnant volume (%) Silent MRA/ FSE MRC	Results
1	0	0.0	4.5 × 3.6 × 4.0	33.9	0.0	complete
2	2.0 × 2.0 × 1.8	3.8	4.8 × 3.9 × 4.8	47.1	8.0	remnant, follow-up
3	3.3 × 2.1 × 3.4	12.3	5.6 × 10.9 × 7.4	236.5	5.2	remnant, follow-up
4	5.6 × 5.5 × 5.7	91.9	13.9 × 14.0 × 17.8	1813.7	5.1	remnant, follow-up
5	1.0 × 1.9 × 2.3	2.3	2.8 × 3.0 × 2.7	11.9	19.3	remnant, follow-up
6	0	0.0	3.7 × 4.1 × 4.5	35.7	0.0	complete
7	3.5 × 2.5 × 3.9	17.9	6.1 × 5.8 × 6.5	120.4	14.8	remnant, follow-up
8	3.4 × 2.2 × 2.8	11.0	4.3 × 3.4 × 3.9	29.9	36.7	remnant, follow-up
9	3.2 × 2.8 × 3.1	14.5	19.9 × 2.8 × 11.4	332.6	4.4	remnant, follow-up
10	5.0 × 4.9 × 4.0	52.4	16.2 × 15.5 × 17.3	2274.5	2.3	remnant, follow-up
11	4.0 × 3.0 × 1.3	8.2	4.5 × 5.2 × 7.9	80.2	6.9	remnant, follow-up
12	10.5 × 7.4 × 7.0	284.8	11.9 × 12.2 × 10.9	828.6	34.4	remnant, follow-up



Case 5: AComA Aneurysm (Pre-Post)



結論

- ① TOF-MRAでは、metal-artifactにより、全例で neck-complexの描出が不完全であった。
- ② ASL(Silent)-MRAは非造影-MRAであり、contrast の副作用なく、管腔内動脈血(血管)形態が表示され、coiled-neckは、metal-artifactなく描出され、neck-remnantの評価が可能であった。
- ③ FSE-MRCでは、coiled-domeが瘤周囲構造物とともに描出された。
- ④ 3D multifusion image (②+③)では、Coil後のneck-remnantとdomeが良好に描出される為、Coil治療後脳動脈瘤の非侵襲的follow-upに有用と考えられた。

ご清聴ありがとうございました。

